



Intercomparison of integrating nephelometers Project No.: IN-2019-1-1

Basic informations:

Location of the quality assurance: TROPOS, Lab 121
Date: 3 June - 7 June 2019

Principal Investigator	Home Institution	Participant	Instrument
J. M. Pichon	LaMP	J. M. Pichon	70511131

1 Intercomparison summary

Status on arrival

No issues due to transportation or other damages.

Zerocheck

The noise level of the instrument is in the normal range. The average noise (1σ) for the all wavelengths was less equal 0.38 Mm^{-1} for full scattering and 0.36 Mm^{-1} for backscattering. The background level was acceptable with deviations of less equal 0.22 Mm^{-1} for full scattering and 0.43 Mm^{-1} for backscattering.

Spancheck

The span check was acceptable with deviations of less equal 8.6%.

Inspection

The instrument was almost clean with just slight contamination. Due to the good results, a recalibration was not performed.

Comparison to reference nephelometer

Before inspection and recalibration

The results from intercomparison to reference device were acceptable with deviations in the range of -3.8% to 3.8% .

After inspection and recalibration

No information can be given.

Recommendations

No recommendations.

Overall assessment

The instrument meets the requirements.

2 Details

Configuration parameters

Zerocheck

Table 1: Noise parameters of nephelometer (SN 70511131) measured with filtered air.

Wavelength [nm]	total scattering		backscattering	
	mean [Mm ⁻¹]	std.dev. [Mm ⁻¹]	mean [Mm ⁻¹]	std.dev. [Mm ⁻¹]
450	0.1	0.38	0.23	0.36
530	0.16	0.25	0.08	0.28
700	0.22	0.22	0.43	0.25

Spancheck

Table 2: Percentage deviation of measured values from nephelometer (SN 70511131) to theoretical values for CO₂

Wavelength [nm]	total scattering	backscattering
	deviation [%]	deviation [%]
450	-6.6	-6.5
530	-6	-8.6
700	-8.5	-5.4

Comparison to reference nephelometer before inspection and recalibration

Table 3: Comparison of nephelometer (SN 70511131) to reference nephelometer Aurora4000 (SN 14-1408) before inspection and recalibration. Testaerosol is ammonium sulfate.

Wavelength [nm]	total scattering slope	R2	backscattering slope	R2
450	1.037	1	0.962	0.999
525	1.038	1	0.964	0.998
635	0.977	1	1.017	0.998




Figure 1: Correlation of scattering coefficients from nephelometer (SN 70511131) and reference nephelometer Aurora4000 (SN 14-1408) before inspection and recalibration. Testaerosol is ammonium sulfate.