



Intercomparison of Integrating Nephelometers

Project No.: AP-2016-2-6

Basic Information:

Location of the quality assurance: TROPOS

Delivery Date: 24 July, 2017

Principal Investigator	Home Institution	Participant	Instrument
Ivo Kalapov	INRE-BAS	I. Kalpov	CLAP, SN 10.022

Intercomparison summary

Status on arrival: The instrument failed the intercomparison because of a hardware error.

Recommendations: Instruments needs to be repaired and registered for another intercomparison workshop.