

## Intercomparison of Condensation Particle Counter

<i>Project No.:</i>	CPC-2017-1-9
<i>Principal Investigator:</i>	Kay Weinhold
<i>Home Institution:</i>	TROPOS
<i>Participant:</i>	Round Robin Tour in ACTRIS (Dp50 cal. to 10nm)
<i>Candidate:</i> <i>Counter (SN):</i>	<b>DE-TROPOS Box4</b> TSI CPC Model 3772; SN: 3772154301
<i>Location of the quality assurance:</i>	TROPOS Leipzig, lab 130
<i>Comparison period:</i>	January 25, 2017
<i>Last Intercomparison (with Project No.):</i>	
<i>TROPOS Reference Instrument:</i>	Electrometer: TSI model 3068B #70838596, Last calibration in April 2016
<i>Additional Equipment:</i>	Bubble flow meter 'Giliblator', Gilian (Sensidyne)

### Summary of Intercomparison


#### *Pre-Status:*

The candidate did not pass the quality standards of ACTRIS and GAW. The candidate reached 84% efficiency at 40 nm. The Dp50 is at 9.95 nm. The CPC efficiency curve does not correspond to the standard. During the Round Robin Tour the total CPC in Box4 "TROPOS SN3772154301" showed status values out of the range. TROPOS decided to ship the instrument back to TROPOS for maintenance before continuing the intercomparison.

#### *Final Status:*

The candidate passed the quality standards of ACTRIS and GAW. It was necessary to clean and recalibrate the candidate. The CPC efficiency curve corresponds to the standard. The candidate reached 99% efficiency at 40 nm. The Dp50 is at 9.82 nm. The Box4 is going back to the last station for a second intercomparison.

## Laboratory setup




**Status of the candidate:**

<i>CPC status</i>	Pre-Status	Final Status
<i>power/status</i>	LED green	LED green
<i>saturator temp</i>	39.0 °C	39.0 °C
<i>condenser temp</i>	23.9 °C	24.0 °C
<i>optics temp</i>	40.0 °C	40.0 °C
<i>cabinet temp</i>	33.2 °C	26.1 °C
<i>ambient pressure</i>	101.2 kPa	101.2 kPa
<i>orifice pressure</i>	85.4 kPa	83.0 kPa
<i>nozzle pressure</i>	? kPa	3.3 kPa
<i>laser current</i>	23 mA	42 mA
<i>liquid level</i>	full	full
<i>Aerosol flow (l/min)</i>	0.985 l/min	1.001 l/min

**Special Information regarding to the Candidate:**

<i>Was it necessary to:</i>	yes/no	information
<i>do a second run</i>	Yes	Problems during the Round Robin Tour. CPC was dirty and needed a service.
<i>clean the optics</i>	no	
<i>clean the nozzle</i>	yes	Black maybe soot
<i>clean the saturator</i>	yes	Black maybe soot
<i>change the wick</i>	no	
<i>change the laser</i>	no	
<i>change internal settings</i>	no	


## CPC efficiency curve of the candidate: Pre-Status



**Figure 01:** CPC efficiency curve. Based on Electrometer TSI 3068B. Serial number: 70838596

## Measured data of the candidate: Pre-Status

$D_p$ in nm	counting efficiency
40 nm	<b>0.84</b>
30 nm	<b>0.83</b>
20 nm	<b>0.78</b>
15 nm	<b>0.70</b>
10 nm	<b>0.47</b>
8 nm	<b>0.27</b>
7 nm	<b>0.10</b>
6 nm	<b>0.00</b>

**CPC efficiency curve of the candidate: Final-Status****Figure 02:** CPC efficiency curve. Based on Electrometer TSI 3068B. Serial number: 70838596**Measured data of the candidate: Final-Status**

$D_p$ in nm	counting efficiency
40 nm	<b>0.98</b>
30 nm	<b>0.98</b>
20 nm	<b>0.91</b>
15 nm	<b>0.81</b>
10 nm	<b>0.52</b>
8 nm	<b>0.29</b>
7 nm	<b>0.12</b>
6 nm	<b>0.00</b>



**Physikalisch-Technische Bundesanstalt**  
**Braunschweig und Berlin**  
Nationales Metrologieinstitut



## Kalibrierschein

Calibration Certificate

Gegenstand:  
Object: Aerosol Elektrometer  
ID-Nummer: TSI\_3068B\_70838596

Hersteller:  
Manufacturer: TSI Incorporated

Typ:  
Type: 3068B

Kennnummer:  
Serial No.: 70838596

Auftraggeber:  
Applicant: AG 3.23 - Herr Nowak, Tel.: 3228

Anzahl der Seiten:  
Number of pages: 4

Geschäftszeichen:  
Reference No.: PTB AG 2.11-1461323527-0327

Kalibrierzeichen:  
Calibration mark: 20790 PTB 16

Datum der Kalibrierung:  
Date of calibration: 22.04.2016

Im Auftrag:  
On behalf of PTB: Braunschweig, 04. Mai. 2016

Im Auftrag:  
On behalf of PTB:

Siegel  
Seal



391 00B n

Dr. B. Schumacher

C. Rohrig

Kalibrierscheine ohne Unterschrift und Siegel haben keine Gültigkeit. Dieser Kalibrierschein darf nur unverändert weiterverbreitet werden. Auszüge bedürfen der Genehmigung der Physikalisch-Technischen Bundesanstalt.

Calibration certificates without signature and seal are not valid. This calibration certificate may not be reproduced other than in full. Extracts may be taken only with permission of the Physikalisch-Technische Bundesanstalt.