



Leibniz Institute for
Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Intercomparison of Condensation Particle Counter

Project No.: CPC-2019-3-27

Principal Investigator: Marcos Andrade

Home Institution: Chacaltaya GAW Station
Carrera de Fisica
Universidad Mayor de San Andres
Campus Universitario Cota-Cota
La Paz, Bolivia

Participant: –

Candidate: CPC Bolivia

Counter (SN): TSI CPC Model 3772 SN70826011

Location of the quality assurance: TROPOS Leipzig, lab 130

Comparison period: August 13, 2019

Last Intercomparison (with Project No.):

TROPOS Reference Instrument: Electrometer: TSI model 3068B
#70838596, Last calibration in September 2018

Additional Equipment: Bubble flow meter 'Gilibrator', Gilian (Sensidyne)
#1711008-S, Last calibration in January 2018

Summary of Intercomparison

Status:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 100% efficiency at 40 nm. The Dp50 is at 7.89 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW. Information: It was necessary to clean the CPC, change the wick and laser. The CPC is calibrated.

Page 1 / 4


Leibniz-Institut für Troposphärenforschung e.V.
Telefon: +49 341 2717-7060
Telefax: +49 341 2717-99-7060
info@tropos.de
<http://www.tropos.de>

Commerzbank Leipzig
KTO 102 14 50
BLZ 860 400 00
IBAN: DE77 8604 0000 0102 1450 00
SWIFT CODE: COBADEFF 860

Mitglied der

Leibniz-Gemeinschaft

Laboratory Setup:





Leibniz Institute for
Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Date of arrival of instrument in calibration lab:

August 13, 2019

Instrument:

Condensation Particle Counter

Model and serial number of instrument:

CPC 3772 S/N 70826011

Result of physical inspection:

no damages

Result of functional test:

cleaning necessary, laser broken

Internal parameters of instrument

nominal flow rate 1.0 l/min

**Model and identification number of
aerosol electrometer:**

TSI Electrometer Model 3068, S/N 70838596

Electrometer calibration certificate:

*September 05, 2018, calibrated at PTB
Braunschweig*

**Corrections of electrometer, for instance,
differing flow rate:**

*Within tolerance range (+/-2%); reference: 4.0
l/min, measured: 4.00 l/min*

Software for recording:

*LabView 2010; National Instruments; Program
„LabCount.vi“*

Date of calibration:

August 13, 2019

Lab temperature and pressure:

23°C, 992.5mbar

Measured aerosol flow rate of CPC:

1.019 l/min

Uncertainty in measured flow rate:

3%

Flowmeter used:

*Gilian Gilibrator V; S/N 1711008-S,
January, 2018*

Particles and gases used for calibration:

silver particles and nitrogen

Method of particle generation:

tube furnace generator

Zero measurement of instrument:

0 particles/cm³ in 5 minutes

Results (using pulse output):

Particle size (nm)	40	30	20	15	12
Number concentration (cm-3)	2436	2347	2383	3024	2226
Counting efficiency η	1.01	1.00	1.00	0.97	0.88
Particle size (nm)	10	09	08	07	06
Number concentration (cm-3)	1353	987	908	423	46
Counting efficiency η	0.77	0.68	0.52	0.31	0.04
Particle size (nm)	40				
Number concentration (cm-3)	1536				
Counting efficiency η	1.02				

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Special Information regarding to the Candidate:

Was it necessary to:	yes/no	information
do a second run	no	-
clean the optics	yes	-
clean the nozzle	yes	-
clean the saturator	yes	-
change the wick	yes	new wick
change the laser	yes	laser old
change internal settings	no	-




Fig. 1: Counting efficiency for CPC Bolivia 3772 SN70826011 against aerosol electrometer 3068 SN 70838596; silver particles between 6 and 40 nm were used for calibration; the calculated D_{p50} is 7.89 nm.

Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB
from display	39	22	40	31.1	99.7
Status	P OR	P NO	Laser	LV	flow
from display	71.2	2.7	56	full	1.019

Date of issue: August 13, 2019

Reviewed: TROPOS / Kay Weinhold

Page 4 / 4