







Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Intercomparison of Condensation Particle Counter

CPC-2019-4-3 Project No.:

Principal Investigator: Jean-François Doussin

Home Institution: CNRS IRCELYON

Participant: Sebastien Perrier

Clement Dubois

Candidate: **IRCELYON**

Counter (SN): TSI CPC Model 3772 #131003

Location of the quality assurance: TROPOS Leipzig, lab 130

September 17, 2019 Comparison period:

Last Intercomparison (with Project No.):

TROPOS Reference Instrument: Electrometer: TSI model 3068B

#70838596, Last calibration in September 2018

Additional Equipment: Bubble flow meter 'Gilibrator', Gilian (Sensidyne)

#1711008-S, Last calibration in January 2018

Summary of Intercomparison

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 100% efficiency at 40 nm. The Dp50 is at 9.38 nm in the final-status. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Info: The candidate passed the pre-status. To increase the performance, TROPOS opened the CPC and checked and cleaned it. It was not necessary to replace or repair anything.

Page 1 / 5

Commerzbank Leipzig








for Aerosol Physics





eibniz Institute for.
Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig



Leibniz-Institut für Troposphärenforschung e.V.
Telefon: +49 341 2717-7060
Telefax: +49 341 2717-99-7060

Telefax: +49 341 2717-99-7060 info@tropos.de http://www.tropos.de Commerzbank Leipzig KTO 102 14 50 BLZ 860 400 00 IBAN: DE77 8604 0000 0102 1450 00 SWIFT CODE: COBADEFF 860

Leibniz-Gemeinschaft









Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Date of arrival of instrument in calibration lab: September 17, 2019

Instrument:

Model and serial number of instrument: CPC 3772 S/N 131003

for Aerosol Physics

Result of physical inspection: no damages **Result of functional test:** cleaned

Internal parameters of instrument nominal flow rate 1.0 l/min

Model and identification number of

aerosol electrometer: TSI Electrometer Model 3068, S/N 70838596

Electrometer calibration certificate: September 05, 2018, calibrated at PTB

Braunschweig

Corrections of electrometer, for instance,

differing flow rate:

l/min, measured: 4.00 l/min

Within tolerance range (+/-2%); reference: 4.0

Condensation Particle Counter

Software for recording: LabView 2010; National Instruments; Program

"LabCount.vi"

Date of calibration: September 17, 2019 Lab temperature and pressure: 23°C, 1004 mbar

Measured aerosol flow rate of CPC: pre-status: 0.989 l/min, final status: 1.024 l/min

Uncertainty in measured flow rate: 3%

Flowmeter used: Gilian Gilibrator V; S/N 1711008-S,

January, 2018

Particles and gases used for calibration: silver particles and nitrogen Method of particle generation: tube furnace generator **Zero measurement of instrument:** 0 particles/cm³ in 5 minutes

Special Information regarding to the Candidate:

Was it necessary to:	yes/no	information
do a second run	yes	dirty – increased flow
clean the optics	yes	cleaned
clean the nozzle	yes	cleaned
clean the saturator	yes	cleaned
change the wick	yes	-
change the laser	no	-
change internal settings	no	-

SWIFT CODE: COBADEFF 860







for Aerosol Physics





Leibniz Institute for Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Results (using pulse output): pre-status

Particle size (nm)	40	30	20	15	12
Number concentration (cm-3)	1142	1562	1640	918	1292
Counting efficiency η	0.98	0.97	0.94	0.87	0.75
Particle size (nm)	10	09	08	07	06
Number concentration (cm-3)	625	690	571	142	1
Counting efficiency η	0.60	0.45	0.25	0.06	0.0
Particle size (nm)					
Number concentration (cm-3)					
Counting efficiency η					




Fig. 1: pre-status: Counting efficiency for TSI CPC Model 3772 #131003 against aerosol electrometer 3068 S/N 70838596; silver particles between 6 and 40 nm were used for calibration; the calculated Dp50 is 9.25 nm.

Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB
from display	39	22	40	32.6	99.8
Status	P OR	P NO	Laser	LV	flow
from display	81.4	2.7	29	full	0.989

Page 4 / 5


http://www.tropos.de

KTO 102 14 50 BLZ 860 400 00 IBAN: DE77 8604 0000 0102 1450 00 SWIFT CODE: COBADEFF 860

Commerzbank Leipzig

Mitglied der Leibniz-Gemeinschaft









Leibniz Institute for Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Results (using pulse output): final-status

Results (using pulse output). Illiai-status						
Particle size (nm)	40	30	20	15	12	
Number concentration (cm-3)	932	-	1464	-	-	
Counting efficiency η	1.01	-	1.00	-	-	
Particle size (nm)	10	09	08	07	06	
Number concentration (cm-3)	848	-	522	-	1	
Counting efficiency η	0.59	-	0.25	-	0.00	
Particle size (nm)						
Number concentration (cm-3)						
Counting efficiency η						

for Aerosol Physics




Fig. 1: final-status: Counting efficiency for TSI CPC Model 3772 #131003 against aerosol electrometer 3068 S/N 70838596; silver particles between 6 and 40 nm were used for calibration; the calculated Dp50 is 9.38 nm.

Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB
from display	39	22	40	32.6	99.8
Status	P OR	P NO	Laser	LV	flow
from display	81.4	2.6	29	full	1.024

Date of issue: September 17, 2019 Reviewed: TROPOS / Kay Weinhold

Page 5 / 5

Leibniz-Gemeinschaft