



Leibniz Institute for
Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Intercomparison of Condensation Particle Counter

<i>Project No.:</i>	CPC-2019-4-4
<i>Principal Investigator:</i>	Jiri Sperka
<i>Home Institution:</i>	Czech Metrology Institute
<i>Participant:</i>	Jiri Sperka
<i>Candidate: Counter (SN):</i>	CMI TSI CPC Model 3772 #3772133801
<i>Location of the quality assurance:</i>	TROPOS Leipzig, lab 130
<i>Comparison period:</i>	September 17, 2019
<i>TROPOS Reference Instrument:</i>	Electrometer: TSI model 3068B #70838596, Last calibration in September 2018
<i>Additional Equipment:</i>	Bubble flow meter 'Gilibrator', Gilian (Sensidyne) #1711008-S, Last calibration in January 2018

Summary of Intercomparison

Status:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 99% efficiency at 40 nm. The Dp50 is at 8.65 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Page 1 / 4

Leibniz-Institut für Troposphärenforschung e.V.
Telefon: +49 341 2717-7060
Telefax: +49 341 2717-99-7060
info@tropos.de
<http://www.tropos.de>


Commerzbank Leipzig
KTO 102 14 50
BLZ 860 400 00
IBAN: DE77 8604 0000 0102 1450 00
SWIFT CODE: COBADEFF 860

Mitglied der

Leibniz-Gemeinschaft

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Laboratory Setup:





Leibniz Institute for
Tropospheric Research

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Date of arrival of instrument in calibration lab: *September 17, 2019*
Instrument: *Condensation Particle Counter*
Model and serial number of instrument: *CPC 3772 S/N 3772133801*

Result of physical inspection: *no damages*
Result of functional test: *no problems*

Internal parameters of instrument *nominal flow rate 1.0 l/min*

Model and identification number of aerosol electrometer: *TSI Electrometer Model 3068, S/N 70838596*

Electrometer calibration certificate: *September 05, 2018, calibrated at PTB Braunschweig*

Corrections of electrometer, for instance, differing flow rate: *Within tolerance range (+/-2%); reference: 4.0 l/min, measured: 4.00 l/min*

Software for recording: *LabView 2010; National Instruments; Program „LabCount.vi“*

Date of calibration: *September 17, 2019*
Lab temperature and pressure: *23°C, 1001 mbar*
Measured aerosol flow rate of CPC: *1.001 l/min*
Uncertainty in measured flow rate: *3%*
Flowmeter used: *Gilian Gilibrator V; S/N 1711008-S, January, 2018*

Particles and gases used for calibration: *silver particles and nitrogen*
Method of particle generation: *tube furnace generator*
Zero measurement of instrument: *0 particles/cm³ in 5 minutes*

Results (using pulse output):

Particle size (nm)	40	30	20	15	12
Number concentration (cm-3)	1140	1600	1707	977	1429
Counting efficiency η	0.98	1.00	0.98	0.92	0.83
Particle size (nm)	10	09	08	07	06
Number concentration (cm-3)	740	929	1021	596	83
Counting efficiency η	0.71	0.61	0.45	0.25	0.04
Particle size (nm)					
Number concentration (cm-3)					
Counting efficiency η					

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig

Special Information regarding to the Candidate:

Was it necessary to:	yes/no	information
do a second run	no	-
clean the optics	no	-
clean the nozzle	no	-
clean the saturator	no	-
change the wick	no	-
change the laser	no	-
change internal settings	no	-




Fig. 1: Counting efficiency for TSI CPC Model 3772 #3772133801 against aerosol electrometer 3068 S/N 70838596; silver particles between 6 and 40 nm were used for calibration; the calculated $Dp50$ is 8.24 nm.

Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB
from display	39	22	40	32.2	99.9
Status	P OR	P NO	Laser	LV	flow
from display	80.3	2.7	41	full	1.001

Date of issue: September 17, 2019

Reviewed: TROPOS / Kay Weinhold

Page 4 / 4