



World Calibration Centre  
for Aerosol Physics



European Center for Aerosol Calibration



Leibniz Institute for  
Tropospheric Research

## Intercomparison of Condensation Particle Counter

**Project No.:** CPC-2020-1-5

**CPC Model:** Airmodus

**CPC Serial Number:** 2331645306

**Principal Investigator:** Dr. Joonas Vanhanen

**Home Institution:** Airmodus, Finland

**Participant:** -

**Description:** Calibration of a Condensation Particle Counter (CPC, Model Airmodus)

**Date of Calibration:** March 18, 2020

### Summary of Intercomparison:

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 100% efficiency at 40 nm. The  $Dp_{50}$  is at 8.70 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Certificate / Reference: WCCAP

Date of issue: March 20, 2020

Signature:

Reviewed by: **TROPOS**

Name: **Kay Weinhold**

Page 1 / 3



World Calibration Centre  
for Aerosol Physics



European Center for Aerosol Calibration



Leibniz Institute for  
Tropospheric Research

**Date of arrival of instrument in calibration lab:** *March 10, 2020*  
**Instrument:** *Condensation Particle Counter*  
**Model and serial number of instrument:** *CPC Airmodus SN2331645306*

**Result of physical inspection:** *no damages*  
**Result of functional test:** *functional test successful, no problems*

**Internal parameters of instrument** *nominal flow rate 1.0 l/min*

**Model and identification number of aerosol electrometer:** *TSI Electrometer Model 3068, S/N 70838596*

**Electrometer calibration certificate:** *September 5, 2018, calibrated at PTB Braunschweig*

**Corrections of electrometer, for instance, differing flow rate:** *Within tolerance range (+/-2%); reference: 4.0 l/min, measured: 4.000 l/min*

**Software for recording:** *LabView 2010; National Instruments; Program „LabCount.vi“*

**Date of calibration:** *March 18, 2020*  
**Lab temperature and pressure:** *23.0°C, 1008 mbar*  
**Measured aerosol flow rate of CPC:** *1.035 l/min*  
**Uncertainty in measured flow rate:** *3%*  
**Flowmeter used:** *Gilian Gilibrator V; S/N 1711008-S, January, 2018*

**Particles and gases used for calibration:** *silver particles and nitrogen*  
**Method of particle generation:** *tube furnace generator*  
**Zero measurement of instrument:** *0 particles/cm<sup>3</sup> in 5 minutes*

**Results (using pulse output and logging via TROPOS LabVIEW software):**

<b>Particle size (nm)</b>	<b>40</b>	<b>30</b>	<b>20</b>	<b>15</b>	<b>10</b>	<b>09</b>
<b>Number concentration (cm-3)</b>	1337	1506	995	1172	1360	991
<b>Counting efficiency η</b>	0.99	1.00	1.00	0.97	0.74	0.52
<b>Particle size (nm)</b>	<b>08</b>	<b>07</b>	<b>06</b>	<b>05</b>	<b>40</b>	
<b>Number concentration (cm-3)</b>	621	249	13	0	1088	
<b>Counting efficiency η</b>	0.33	0.11	0.01	0	1.00	



World Calibration Centre  
for Aerosol Physics



European Center for Aerosol Calibration



Leibniz Institute for  
Tropospheric Research




Fig. 1: Counting efficiency for CPC Airmodus S/N 2331645306 against aerosol electrometer 3068 S/N 70838596; silver particles between 5 and 40 nm were used for calibration; the calculated  $Dp_{50}$  is 8.70 nm.

**Status information:**

<b>Status</b>	<b>T SAT</b>	<b>T CON</b>	<b>T OPT</b>	<b>T CAB</b>	<b>P AMB</b>	<b>P VAC</b>
from display	35	20	36		102	-
<b>Status</b>	<b>P OR</b>	<b>P NO</b>	<b>Laser</b>	<b>LV</b>	<b>flow</b>	<b>P INLET</b>
from display	81	1	53	full	1.035	-

**Date of issue:** March 20, 2019

Reference: TSI electrometer, model 3068, SN 70838596

Reviewed: TROPOS / Kay Weinhold

Page 3 / 3