



**World Calibration Centre  
for Aerosol Physics**

Leibniz-Institut für Troposphärenforschung Permoserstraße 15 04318 Leipzig



**Leibniz Institute for  
Tropospheric Research**

**CPC Model:** TSI CPC 3750

**CPC Serial Number:** 3750201202

**Customer:** TSI Instruments Ltd.

**Description:** Calibration of a Condensation Particle Counter (CPC, Model 3750)

**Date of Calibration:** September 16, 2020

**Summary of Intercomparison:**

The candidate passed the quality standards of ACTRIS and GAW. The candidate reached 99% efficiency at 40 nm. The Dp50 is at 6.92 nm. The CPC efficiency curve corresponds to the standard of ACTRIS and GAW.

Certificate / Reference: WCCAP

Date of issue: September 16, 2020      Signature:

Reviewed by: **TROPOS**

Name: **Kay Weinhold**

Page 1 / 4



World Calibration Centre  
for Aerosol Physics



Leibniz Institute for  
Tropospheric Research

**Date of arrival of instrument in calibration lab:** *September, 2020*  
**Instrument:** *Condensation Particle Counter*  
**Model and serial number of instrument:** *CPC 3750 S/N 3750201202*

**Result of physical inspection:** *no damages*  
**Result of functional test:** *functional test successful, no problems*

**Internal parameters of instrument** *nominal flow rate 1.0 l/min*

**Model and identification number of aerosol electrometer:** *TSI Electrometer Model 3068, S/N 70838596*

**Electrometer calibration certificate:** *September 5, 2018, calibrated at PTB Braunschweig*

**Corrections of electrometer, for instance, differing flow rate:** *Within tolerance range (+/-2%); reference: 4.0 l/min, measured: 4.000 l/min*

**Software for recording:** *LabView 2010; National Instruments; Program „LabCount.vi“*

**Date of calibration:** *September 16, 2020*  
**Lab temperature and pressure:** *23.0°C, 1001 mbar*  
**Measured aerosol flow rate of CPC:** *0.992 l/min*  
**Uncertainty in measured flow rate:** *3%*  
**Flowmeter used:** *Gilian Gilibrator V; S/N 1711008-S, January, 2018*

**Particles and gases used for calibration:** *silver particles and nitrogen*  
**Method of particle generation:** *tube furnace generator*  
**Zero measurement of instrument:** *0 particles/cm<sup>3</sup> in 5 minutes*

**Results (using pulse output and logging via TROPOS Labview software):**

Particle size (nm)	40	30	20	15	14	11	10
Number concentration (cm-3)	1288	1360	1770	1071	1459	826	1467
Counting efficiency $\eta$	0.99	1.00	0.99	0.96	0.95	0.90	0.84
Particle size (nm)	09	08	07	06	05	40	
Number concentration (cm-3)	1399	799	1090	604	23	1345	
Counting efficiency $\eta$	0.77	0.66	0.50	0.28	0.02	0.99	



World Calibration Centre  
for Aerosol Physics



Leibniz Institute for  
Tropospheric Research




Fig. 1: Counting efficiency for CPC 3750 S/N 3750201202 against aerosol electrometer 3068 S/N 70838596; silver particles between 5 and 40 nm were used for calibration; the calculated  $D_{p50}$  by the BNC Pulse Output on Labview is 6.92 nm.

#### Status information:

Status	T SAT	T CON	T OPT	T CAB	P AMB	P VAC
from display	39.0	19.8	40.0	22.4	100.9	82.8
Status	P OR	P NO	Laser	LV	flow	P INLET
from display	81.9	2.45	41	full	0.992	-0.3

#### Results:

using pulse output and logging via TROPOS Labview software: without coincidence correction

Concentration EM in #/cm <sup>3</sup>	61792	50032	32548	28098	17937
Number concentration without coincidence correction (cm-3)	47550	40065	27719	24300	16225
Counting efficiency $\eta$	0.77	0.80	0.85	0.86	0.90
Concentration EM in #/cm <sup>3</sup>	12485	5332	1436		
Number concentration without coincidence correction (cm-3)	11536	5046	1401		
Counting efficiency $\eta$	0.92	0.95	0.98		



World Calibration Centre  
for Aerosol Physics



Leibniz Institute for  
Tropospheric Research

## Results:

using USB-C connection and logging via TSI software: with coincidence correction					
Concentration EM in $\#/cm^3$	61792	50032	32548	28098	17937
Number concentration with coincidence correction ( $cm^{-3}$ )	60128	48849	31927	27558	17820
Counting efficiency $\eta$	0.97	0.98	0.98	0.98	0.99
Concentration EM in $\#/cm^3$	12485	5332	1436		
Number concentration with coincidence correction ( $cm^{-3}$ )	12497	5345	1463		
Counting efficiency $\eta$	1.00	1.00	1.02		




Fig. 2: Linearity test for TSI CPC 3750 SN 3750201202 against aerosol electrometer 3068 SN 70838596; silver particles with a diameter of 30 nm were used for number concentrations between 1000 and 60000 particles per  $cm^3$ .

Date of issue: September 16, 2020

Reference: TSI electrometer, model 3068, SN 70838596

Reviewed: TROPOS / Kay Weinhold