

Intercomparison of Mobility Particle Size Spectrometers

Project No.: MPSS-2017-3-7

Principal Investigator: Harald Flentje

Home Institution: DWD
Albin-Schweiger-Weg 10
82383 Hohenpeißenberg

Participant: Björn Briel

Candidate: **DE-DWD Hohenpeißenberg**

Made by:

Counter (SN): TSI CPC Model 3772, SN: 70711210

Software: TSI

Location of the quality assurance: TROPOS Leipzig, lab 118

Comparison period: May 16, 2017 – May 19, 2017

Last Intercomparison (with Project No.):

Summary of Intercomparison:*Pre-Status:*

The instrument arrived with participant. During the Pre-Status, the performance of the system showed a concentration 1% lower than the TROPOS Reference Instrument No.1. The PSL check showed a correct peak at 201.83 nm.

Final-Status:


During the Final-Status, the performance of the system showed a concentration 2% higher than the TROPOS Reference Instrument No.1. The candidate used the recalibrated TSI CPC model 3772 and their own Kr.85 source. The candidate passed the quality standards of ACTRIS and GAW.

Information about the instruments:

Date of check: May 15, 2017

List of Components	TROPOS Reference MPSS No.1	TROPOS Reference MPSS No.	Candidate
Position	Line 1	-	Line 1
Company	TROPOS	-	TROPOS
Software	TROPOS	-	TROPOS V6.66
CPC-MPSS	TSI CPC, Model 3772	-	TSI CPC, Model 3772
CPC-total	TSI CPC, Model 3010	-	TSI CPC, Model 3772
flow ratio	1.0 : 5.0	-	1.0 : 5.0
source	Kr85	-	Kr85
HV power supply	Positive	-	positive
DMA	Hauke medium	-	Hauke medium
aerosol dryer	✓	-	✓
aerosol RH- sensor	✓	-	✓
aerosol T-sensor	✓	-	✓
sheath RH-sensor	✓	-	✓
sheath T-sensor	✓	-	✓
Sheath dryer	✓	-	✓
pressure sensor	✓	-	✓
info			

Laboratory setup:



Status of the instruments:

Date of check (Pre-Status): May 15, 2017

<i>CPC status</i>	MPSS		Total CPC	
<i>power/status</i>	LED green	-	-	-
<i>saturator temp</i>	39.0	°C	-	°C
<i>condenser temp</i>	22.0	°C	-	°C
<i>optics temp</i>	40.0	°C	-	°C
<i>cabinet temp</i>	30.9	°C	-	°C
<i>ambient pressure</i>	100.8	kPa	-	kPa
<i>orifice pressure</i>	78.3	kPa	-	kPa
<i>nozzle pressure</i>	2.9	kPa	-	kPa
<i>laser current</i>	53	mA	-	mA
<i>liquid level</i>	full	-	-	-

Date of check (Final-Status): May 17, 2017

<i>CPC status</i>	MPSS		Total CPC	
<i>power/status</i>	LED green	-	-	-
<i>saturator temp</i>	39.0	°C	-	°C
<i>condenser temp</i>	22.0	°C	-	°C
<i>optics temp</i>	40.0	°C	-	°C
<i>cabinet temp</i>	30.6	°C	-	°C
<i>ambient pressure</i>	100.1	kPa	-	kPa
<i>orifice pressure</i>	78.1	kPa	-	kPa
<i>nozzle pressure</i>	2.9	kPa	-	kPa
<i>laser current</i>	53	mA	-	mA
<i>liquid level</i>	full	-	-	-

Date of system checks:

<i>date</i>	15.05.2017	17.05.2017			unit
<i>total CPC flow</i>	-	-			l/min
<i>aerosol flow (DMA)</i>	-	-			l/min
<i>aerosol flow (UDMA)</i>	-	-			l/min
<i>aerosol flow (total)</i>	1.017	1.008			l/min
<i>Zero MPSS</i>	0	0			#/cm ³
<i>Zero total CPC</i>	-	-			#/cm ³
<i>PSL 203 nm</i>	201.83	202.5			nm
<i>HV – 0 V</i>	0.007	0.003			V
<i>HV – 5 V</i>	5.0	4.9			V
<i>HV – 100 V</i>	99.67	99.7			V
<i>HV – 1000 V</i>	999.3	999.6			V

Special Information regarding the Candidate:

<i>Was it necessary to:</i>	yes/no (date)	old part (ID/SN)	new part (ID/SN)	information
<i>clean the aerosol inlet</i>	yes			
<i>change aerosol Nafion dryer</i>	no	MT 052913-05-7	-	-
<i>change sheath Nafion dryer</i>	Yes	ND 0.7- 98b	ND 0.7 – 72d	No leak but the drying efficiency is not good
<i>check source</i>	yes	-	-	77-0497 28.5 nSv/h - okay
<i>change HV power supply</i>	no	-	-	
<i>clean/change DMA</i>	yes	-	-	Cleaned
<i>change aerosol RH/T-sensor</i>	no	-	-	
<i>change sheath RH/T-sensor</i>	no	-	-	
<i>change pressure sensor</i>	no	-	-	
<i>change inlet Nafion dryer</i>	no	-	-	-
<i>Total filter</i>	yes	-	-	Sheath air before DMA

PSL Scan and calibration: Latex 203 nm +/- 4 nm




Figure 01: Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on May 15th, 2017.




Figure 02: Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on May 15th, 2017.

Pre-Status of the Candidate: Particle Number Size Distribution




Figure 03: Comparison of mean particle number size distribution of TROPOS Reference MPSS No.1 against DE-DWD Hohenpeißenberg from May 15, 2017 08:00 PM – May 16, 2017 06:00 AM. Multiple charge correction, internal diffusion losses and CPC efficiency are included.

Pre-Status of the Candidate: Time Series




Figure 05: Time series (May 15, 2017 08:00 PM – May 16, 2017 06:00 AM) of the integrated particle number concentration ($N_{10-800nm}$) of the MPSS and total number concentration (N_{total}) of the Reference TSI-CPC Model 3010. The inversion for the

candidate was performed using TSI software. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.

Pre-Status of the Candidate: Correlation




Figure 06: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 SN: 2337 and TROPOS Reference MPSS No.1. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 07: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 SN: 2337 and DE-DWD Hohenpeißenberg. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 08: Linear regression between the number concentrations of the TROPOS Reference MPSS No.1 and DE-DWD Hohenpeißenberg. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.

Final-Status of the Candidate: Particle Number Size Distribution




Figure 09: Comparison of mean particle number size distribution of TROPOS Reference MPSS No.1 against DE-DWD Hohenpeißenberg from May 17, 2017 06:00 PM – May 18, 2017 06:00 AM. Multiple charge correction, internal diffusion losses and CPC efficiency are included.

Final-Status of the Candidate: Time Series




Figure 10: Time series (May 17, 2017 06:00 PM – May 18, 2017 06:00 AM) of the integrated particle number concentration ($N_{10-800\text{nm}}$) of the MPSS and total number concentration (N_{total}) of the Reference TSI-CPC Model 3010. The inversion for the candidate was performed using TSI software. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.

Final-Status of the Candidate: Correlation




Figure 11: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 SN: 2337 and TROPOS Reference MPSS No.1. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 12: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 SN: 2337 and DE-DWD Hohenpeißenberg. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 13: Linear regression between the number concentrations of the TROPOS Reference MPSS No.1 and DE-DWD Hohenpeißenberg. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.