

## Intercomparison of Mobility Particle Size Spectrometers

*Project No.:* MPSS-2018-3-6

*Principal Investigator:* Noemí Pérez

*Home Institution:* Institute of Environmental Assessment and Water Research (IDAEA)  
Consejo Superior de Investigaciones Científicas (CSIC)  
C/ Jordi Girona 18-26  
08034 Barcelona, Spain

*Participant:* -

*Candidate:* **Barcelona**  
*Made by:* **TROPOS**  
*Counter (SN):* **TSI 3772 SN: 3772140203**  
*Software:* **TROPOS Software V6.68**

*Location of the quality assurance:* TROPOS Leipzig, lab 118

*Comparison period:* May 14, 2018 – June 4, 2018

*Last Intercomparison (with Project No.):*

## Summary of Intercomparison:

### Status:


The instrument arrived without participant. The system was updated to the newest version of TROPOS-MPSS. We used all components (e.g., sensor, blower, etc.) from the old instrument. The instrument is running with the TROPOS Software Version 6.68. During the intercomparison, the candidate showed a concentration of 1% above the TROPOS Reference MPSS No.1. The PSL check showed a peak at 204.36 nm. The candidate passed the quality standards of ACTRIS and GAW.

## Information about the instruments:

Date of check: May 11, 2018

<i>List of Components</i>	TROPOS Reference MPSS No.1	Candidate
<i>Position</i>	Line 1.1	Line 2.1
<i>Company</i>	TROPOS	TROPOS
<i>Software</i>	TROPOS V6.68	TROPOS V6.68
<i>CPC-MPSS</i>	TSI CPC, Model 3772	TSI CPC, Model 3772
<i>CPC-total</i>	TSI CPC, Model 3010	-
<i>flow ratio</i>	1.0 : 5.0	1.0 : 5.0
<i>source</i>	Kr.85	Ni.63
<i>HV power supply</i>	Positive	Positive
<i>DMA</i>	Hauke medium	Hauke medium
<i>aerosol dryer</i>	✓	✓
<i>aerosol RH-sensor</i>	✓	✓
<i>aerosol T-sensor</i>	✓	✓
<i>sheath RH-sensor</i>	✓	✓
<i>sheath T-sensor</i>	✓	✓
<i>Sheath dryer</i>	✓	✓
<i>pressure sensor</i>	✓	✓
<i>info</i>		

## Laboratory setup:



### Additional Equipment:

- Bubble flow meter 'Giliblator', Gilian (Sensidyne)
- Thermo Scientific Nanosphere Size Standard PSL 203nm ( $\pm 4$ nm)
- Aerosol nebulizer for PSL (homemade TROPOS)
- Voltcraft multimeter (0-1000V), Keysight Technologies


**Status of the instruments:****Date of system checks:**

<i>date</i>	28.05.2018	unit
<i>total CPC flow</i>	-	l/min
<i>aerosol flow (DMA)</i>	1.028	l/min
<i>aerosol flow (UDMA)</i>	-	l/min
<i>aerosol flow (total)</i>	1.028	l/min
<i>Zero MPSS</i>	0	#/cm <sup>3</sup>
<i>Zero total CPC</i>	-	#/cm <sup>3</sup>
<i>PSL 203 nm</i>	204.36	nm


**Special Information regarding the Candidate:**

<i>Was it necessary to:</i>	yes/no (date)	old part (ID/SN)	new part (ID/SN)	information
<i>clean the aerosol inlet</i>	No	-	-	-
<i>change aerosol Nafion dryer</i>	No	-	-	checked
<i>change sheath Nafion dryer</i>	No	-	-	checked
<i>check source</i>	No	-	-	Used TROPOS Ni.63
<i>change HV power supply</i>	No	-	-	-
<i>clean/change DMA</i>	No	-	-	Checked/cleaned
<i>change aerosol RH/T-sensor</i>	No	-	-	checked
<i>change sheath RH/T-sensor</i>	No	-	-	checked
<i>change pressure sensor</i>	No	-	-	checked
<i>change inlet Nafion dryer (500)</i>	No	-	-	-
<i>Change Total filter</i>	No	-	-	-
<i>NI-card</i>	No	-	-	-

## PSL Scan and calibration: Latex 203 nm +/- 4 nm




**Figure 01:** Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on May 14th, 2018




**Figure 02:** Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on May 14<sup>th</sup>, 2018.

## Status of the Candidate: Particle Number Size Distribution




**Figure 03:** Comparison of mean particle number size distribution of TROPOS Reference MPSS No.1 against MPSS Barcelona from May 14, 2018 20:00 – June 04, 2018 06:00. Multiple charge correction, internal diffusion losses and CPC efficiency are included.

## Status of the Candidate: Time Series and Correlation




**Figure 04:** Time series (May 14, 2018 20:00 – June 04, 2018 06:00) of the integrated particle number concentration ( $N_{10,800nm}$ ) of the MPSS and total number concentration ( $N_{total}$ ) of the Reference TSI-CPC Model 3010. The inversion and correction for the candidate were performed using TROPOS software. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.

b



**Figure 05:** Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3010 Ref 1 and MPSS Barcelona (May 14, 2018 20:00 – June 04, 2018 06:00). All corrections are included.



**Figure 06:** Linear regression between the number concentrations of the TROPOS Reference MPSS No.1 and MPSS Barcelona (May 14, 2018 20:00 – June 04, 2018 06:00). All corrections are included.