

Intercomparison of Mobility Particle Size Spectrometers

Project No.: **MPSS-2020-2-1**

Principal Investigator: **Maik Merkel**

Home Institution: **TROPOS**
Permoserstraße 15
04318 Leipzig

Participant: **Maik Merkel**

Candidate: **DE-TROPOS Chemie**

Made by: **TROPOS**

Counter (SN): **TSI CPC Model 3010, SN: 2006**

Software: **LabView version 6.66**

Location of the quality assurance: **TROPOS Leipzig, lab 118**

Comparison period: **Mar 03, 2020 – Mar 05, 2020**

Last Intercomparison (with Project No.):

Summary of Intercomparison:

Final-Status:


During the final run, the performance of the system showed a concentration 4% higher in comparison to the TROPOS Reference Instrument No.1. The candidate used the TSI CPC model 3010 and their own Kr.85 source. The candidate passed the quality standards of ACTRIS and GAW.

Information about the instruments:

Date of check: Mar 03, 2020

List of Components	TROPOS Reference MPSS No.1	TROPOS Reference MPSS No.	Candidate
<i>Position</i>	Line 1	-	Line 1
<i>Company</i>	TROPOS	-	TROPOS
<i>Software</i>	TROPOS	-	TROPOS V6.66
<i>CPC-MPSS</i>	TSI CPC, Model 3772	-	TSI CPC, Model 3010
<i>CPC-total</i>	TSI CPC, Model 3772	-	-
<i>flow ratio</i>	1.0 : 5.0	-	1.0 : 5.0
<i>source</i>	Ni63	-	Kr85
<i>HV power supply</i>	Positive	-	positive
<i>DMA</i>	Hauke medium	-	Hauke medium
<i>aerosol dryer</i>	✓	-	✓
<i>aerosol RH- sensor</i>	✓	-	✓
<i>aerosol T-sensor</i>	✓	-	✓
<i>sheath RH-sensor</i>	✓	-	✓
<i>sheath T-sensor</i>	✓	-	✓
<i>Sheath dryer</i>	✓	-	✓
<i>pressure sensor</i>	✓	-	✓
<i>info</i>			

Laboratory setup:



Status of the instruments:

Date of check (Pre-Status):

<i>CPC status</i>	MPSS		Total CPC	
<i>power/status</i>	-	-	-	-
<i>saturator temp</i>	-	°C	-	°C
<i>condenser temp</i>	-	°C	-	°C
<i>optics temp</i>	-	°C	-	°C
<i>cabinet temp</i>	-	°C	-	°C
<i>ambient pressure</i>	-	kPa	-	kPa
<i>orifice pressure</i>	-	kPa	-	kPa
<i>nozzle pressure</i>	-	kPa	-	kPa
<i>laser current</i>	-	mA	-	mA
<i>liquid level</i>	-	-	-	-

Date of check (Final-Status):

<i>CPC status</i>	MPSS Reference 1		Total CPC	
<i>power/status</i>	green	-	-	-
<i>saturator temp</i>	39.0	°C	-	°C
<i>condenser temp</i>	22.0	°C	-	°C
<i>optics temp</i>	40.1	°C	-	°C
<i>cabinet temp</i>	29.8	°C	-	°C
<i>ambient pressure</i>	99.1	kPa	-	kPa
<i>orifice pressure</i>	78.2	kPa	-	kPa
<i>nozzle pressure</i>	2.7	kPa	-	kPa
<i>laser current</i>	51	mA	-	mA
<i>liquid level</i>	full	-	-	-

Date of system checks:

<i>date</i>	03.03.2020 Reference 1	03.03.2020 Chemie			unit
<i>total CPC flow</i>	1.031	-			l/min
<i>aerosol flow (DMA)</i>	-	-			l/min
<i>aerosol flow (UDMA)</i>	-	-			l/min
<i>aerosol flow (total)</i>	0.986	1.031			l/min
<i>Zero MPSS</i>	0	0			#/cm ³
<i>Zero total CPC</i>	0	-			#/cm ³
<i>PSL 203 nm</i>	202.70	202.22			nm
<i>HV – 0 V</i>	0.0	6.3 => 0.1			V
<i>HV – 5 V</i>	5.0	11.4 => 4.9			V
<i>HV – 100 V</i>	100.0	106.3 => 100.2			V
<i>HV – 1000 V</i>	1000.0	1007.1 => 1001.3			V

Special Information regarding the Candidate:

<i>Was it necessary to:</i>	yes/no (date)	old part (ID/SN)	new part (ID/SN)	information
<i>clean the aerosol inlet</i>	no			
<i>change aerosol Nafion dryer</i>	no	-	MT 112916-01-2	-
<i>change sheath Nafion dryer</i>	no	-	ND 0.7- 146d	-
<i>check source</i>	no	-	-	-
<i>change HV power supply</i>	no	-	-	-
<i>clean/change DMA</i>	no	-	-	-
<i>change aerosol RH/T- sensor</i>	no	-	-	-
<i>change sheath RH/T- sensor</i>	no	-	-	-
<i>change pressure sensor</i>	no	-	-	-
<i>change inlet Nafion dryer</i>	no	-	-	-
<i>Total filter</i>	no	-	-	-

PSL Scan and calibration: Latex 203 nm +/- 4 nm




Figure 01: Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on Feb 28th, 2020 for the MPSS Reference 1 instrument.




Figure 02: Measurement of latex 203 nm: Particle size distribution (raw concentration) for latex 203 nm on March 3rd, 2020 for the MPSS Chemie.

Final Status of the Candidate: Particle Number Size Distribution




Figure 03: Comparison of mean particle number size distribution of TROPOS MPSS Chemie against DE-TROPOS MPSS Reference No.1 from Mar 03, 2020 4:00 PM – Mar 05, 2020 6:00 AM. Multiple charge correction, internal diffusion losses and CPC efficiency are included.

Final Status of the Candidate: Time Series




Figure 04: Time series (Mar 03, 2020 4:00 PM – Mar 05, 2020 6:00 AM) of the integrated particle number concentration ($N_{10-800nm}$) of the MPSS and total number concentration (N_{total}) of the Reference TSI-CPC Model 3772. The inversion for the candidate was performed using TROPOS software. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.

Final Status of the Candidate: Correlation




Figure 05: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3772 SN: 3772154301 and TROPOS Reference MPSS No.1. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 06: Linear regression between the number concentrations of the TROPOS Reference TSI CPC Model 3772 SN: 3772154301 and DE-TROPOS MPSS Chemie. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.




Figure 07: Linear regression between the number concentrations of the TROPOS Reference MPSS No.1 and DE-TROPOS Chemie. Multiple charge correction, internal diffusion losses and CPC flow corrections are included.