

Intercomparison of Mobility Particle Size Spectrometers

Project No.: MPSS-2022-4-2

Participant: TROPOS MPSS – UBA Schauinsland

Software TROPOS: V7.0
Classifier Model: TROPOS
Classifier HV Power Supply: Positive
Neutralizer Model: Kr85
Impactor Model: none
DMA Model: TROPOS
Detector Model: TSI 3772
Detector Model SN: 3773164005
Detector Model Dp50: 10nm
Manuf. Date CPC: -
Firmware: -

Location of the quality assurance: TROPOS Leipzig, WCCAP

Comparison period: May 02, 2022 – May 06, 2022

Summary of Intercomparison:

The TROPOS MPSS UBA Schauinsland participated in the WCCAP workshop in May 2022. The candidate showed a PSL peak at 202.2 nm. The candidate used the TSI CPC model 3772.

Date of arrival of instrument in calibration lab: May 02, 2022

Instrument: Size Spectrometer TROPOS

Model and serial number of instrument: TROPOS MPSS

Result of physical inspection: no damages

Result of functional test: functional test successful, no problems

Internal parameters of instrument: nominal flow rate 1.0 l/min

Model and identification number of TROPOS Reference MPSS: TROPOS MPSS (positive HV)

Date of calibration: May 02-06, 2022

Lab temperature and pressure: 22.0°C, 1003 mbar

Measured aerosol flow rate of CPC: 0.97 l/min

Uncertainty in measured flow rate: 3%

Flowmeter used: Gilian Gilibrator 3; Basis: 21181001005, cell:21191010004,20491011010, 21191012002; May, 2021

Particles and gases used for calibration: ambient aerosol

Zero measurement of instrument: 0 particles/cm³ in 10 minutes

| | Unit | Status |
|---------------------------|------------|------------|
| Model | - | TSI 3772 |
| SN | - | 3772164005 |
| Firmware | - | 2.16 |
| Date | - | 2022 |
| last service date | - | - |
| Saturator Temperature | °C | 39 |
| Condenser Temperature | °C | 24.2 |
| Optics Temperature | °C | 40 |
| Cabinet Temperature | °C | 34.3 |
| Ambient Pressure | kPa | 100.2 |
| Vacuum Pressure | kPa | - |
| Inlet Pressure | kPa | - |
| Critical Orifice Pressure | hPa | 78.2 |
| Aerosol Nozzle Pressure | kPa | 0.4 |
| Laser Current | mA | 48 |
| Liquid Level | - | full |
| Aerosol Flow (Gili) | l/min | 0.97 |
| Internal Aerosol Flow | l/min | - |
| Zero | avg 10 min | 0 |

PSL Scan: Latex 203 nm +/- 4 nm




Figure 01: Measurement of latex 203 nm – TROPOS MPSS: Particle size distribution of latex 203 nm on May 02th, 2022. The peak shows at 202.2nm.

Intercomparison between TROPOS Reference MPSS and MPSS Schauinsland

03.05.2022 06:00 PM – 04.05.2022 06:00 PM




Figure 02: Time series (May 03, 2022 06 PM – May 04, 2022 06 AM) of the integrated particle number concentration ($N_{10-800nm}$) of the MPSS and total number concentration (N_{total}) of the Reference TSI-CPC Model 3750. Multiple charge correction, internal diffusion losses, CPC flow corrections.




Figure 03: Particle size distribution for TROPOS Reference MPSS and MPSS Schauinsland, flow corrections, multiple charge correction and diffusion loss corrections are included.




Figure 04: Linear regression between TROPOS Reference MPSS and MPSS Schauinsland.




Figure 05: Linear regression between TROPOS Reference T-CPC and TROPOS Reference MPSS.

Date of issue: May, 2022

Reviewed: TROPOS / WCCAP