

## Intercomparison of Mobility Particle Size Spectrometers

*Project No.:* MPSS-2022-WCCAP-111

*Principal Investigator:* Adam Kristensson

*Home Institution:* Lund University

*Participant:* -

*Candidate:*

*Made by:* TROPOS

*Counter (SN):* TSI CPC Model 3772, SN: 3772160801 (2016)  
TSI CPC Model 3756, SN: 3756214301 (2021)

*Software:* TROPOS TSMPS 8.0  
Total

*CPC:* -

*Location of the quality assurance:* TROPOS Leipzig, lab 118

*Comparison period:* February 08, 2022 – February 16, 2022

*Last Intercomparison (with Project No.):* -

### Summary of Intercomparison

*Status:*

The candidate passed the quality standards of ACTRIS and GAW during the intercomparison. The system is within the range of +/-10% of the TROPOS Reference MPSS.

The candidate was in a good status. This instrument is new. It was not necessary to change or repair parts of the inlet, instrument or counter. The zero, high voltage and PSL checks are in the correct range of tolerance.

- 1 -

**Information about the instruments:**

Date of check: 08.02.2022

<i>List of Components</i>	TROPOS Reference MPSS No.1	Lund Dual-MPSS
<i>Position</i>	-	-
<i>Company</i>	TROPOS	TROPOS
<i>Software</i>	TROPOS	TROPOS
<i>CPC-MPSS</i>	TSI CPC, Model 3750	TSI CPC, Model 3772 TSI CPC Model 3756
<i>CPC-total</i>	TSI CPC, Model 3750	-
<i>flow ratio</i>	1.0 : 5.0	1.0 : 5.0 15:1.5
<i>source</i>	Ni-63	Kr85
<i>HV power supply</i>	positive	Positive
<i>DMA</i>	Hauke medium	Hauke medium Hauke short
<i>aerosol dryer</i>	√	√
<i>aerosol RH- sensor</i>	√	√
<i>aerosol T-sensor</i>	√	√
<i>sheath RH-sensor</i>	√	√
<i>sheath T-sensor</i>	√	√
<i>Sheath dryer</i>	√	√
<i>pressure sensor</i>	√	√


Date of check: 08.02.2022

<i>CPC status</i>	TROPOS-MPSS	TROPOS-total CPC	Lund-UCPC	Lund-CPC
<i>power/status</i>	LED green	LED green	LED blue	LED green
<i>saturator temp</i>	39 °C	39 °C	39.0	39 °C
<i>condenser temp</i>	24.1 °C	23.6 °C	10.0	22.0 °C
<i>optics temp</i>	40 °C	40 °C	40.0	40.0 °C
<i>cabinet temp</i>	24.3 °C	22.7 °C	27.3	35.1 °C
<i>ambient pressure</i>	101.7 kPa	101.9 kPa	102.5 kPa	102.5 kPa
<i>orifice pressure</i>	78.0 kPa	78.3 kPa	54.8 kPa	71.6 kPa
<i>nozzle pressure</i>	2.37 kPa	2.54 kPa	-0.3 kPa	Sensor defect
<i>laser current</i>	39 mA	42 mA	44 mA	39 mA
<i>liquid level</i>	full	full	full	full


- 2 -

Date of check: 08.02.2022


	TROPOS Reference MPSS		Lund MPSS	
<i>date</i>	<b>pre- status</b>	<b>final- status</b>	<b>pre- status</b>	<b>final- status</b>
<i>aerosol flow (total)</i>	-	0.98 l/min	-	0.97 l/min
<i>zero</i>	-	0 #/cm <sup>3</sup>	-	0 #/cm <sup>3</sup>
<i>HV – 0 V</i>	-	0 V	-	0 V
<i>HV – 4 mV</i>	-	4.91 V	-	5.1 V
<i>HV – 80 mV</i>	-	99.9 V	-	99.7 V
<i>HV – 800 mV</i>	-	999.9 V	-	1000.1 V
	Lund nano MPSS			
<i>date</i>	<b>pre- status</b>	<b>final- status</b>		
<i>aerosol flow (total)</i>	-	1.514 l/min		
<i>zero</i>	-	0 #/cm <sup>3</sup>		
<i>HV – 0 V</i>	-	5.3 V		
<i>HV – 15 mV</i>	-	17.5 V		
<i>HV – 50 mV</i>	-	99.7 V		
<i>HV – 100 mV</i>	-	35.1 V		
<i>HV – 1000 mV</i>	-	350.3 V		




**Figure 01:** Measurement of latex 203 nm Lund MPSS: Raw particle number size distribution of latex 203 nm on February 8<sup>th</sup> 2022. The peak shows at 203.6nm.




**Figure 02:** Measurement of 20 nm silver particles with the Lund Nano-MPSS: Particle size distribution on February 8<sup>th</sup> 2022. The peak shows at 20.7 nm.




**Figure 03:** Time series (8.02.2022 09:00 AM – 9.02.2022 09:00 AM) of the integrated particle number concentration ( $N_{10-800}$ ) of the MPSS of the Lund Dual-MPSS system (red), integral particle number concentration ( $N_{10-800}$ ) of the Reference MPSS 1 (black) and number concentration measured by the total CPC 3750 SN 3750200901 (blue). The gray shaded area represents  $\pm 10$  of the number concentration measured by the total CPC. Multiple charge correction, internal diffusion losses, CPC flow corrections.




**Figure 04:** Raw particle number size distribution for TROPOS Reference MPSS No.1 (black) and Lund Dual-MPSS (green: Nano-MPSS; red: MPSS). Raw data from the instruments without any correction applied.




**Figure 05:** Raw particle number size distribution for TROPOS Reference MPSS No.1 (black) and Lund Dual-MPSS (green: Nano-MPSS; red: MPSS). The shaded area represents the 10% error margin of the TROPOS Reference MPSS No.1. Transfer function correction applied for the different aerosol to sheath air ratios (Nano-MPSS 1.5:15 and MPSS 1:5) resulting in a reduction of the MPSS concentrations by a factor of 2. The raw concentration of the Nano-MPSS has been adjusted to the MPSS at 40 nm. This is necessary due to uncertainty of the nominal flow rate of the UCPC of 50 cc/min (+/- 20%).




**Figure 06:** Time series of the overlap correction applied to the UCPC of the Lund Nano-MPSS in figure 5.




**Figure 07:** Inverted particle number size distributions (8.2.2022 – 9.2.2022) for the TROPOS reference MPSS 1 (black) and the Lund Dual-MPSS system (red). The shaded area represents the 10% error margin of the TROPOS Reference MPSS No.1. Flow corrections, multiple charge correction and diffusion loss corrections are included.



**Figure 08:** Inverted particle number size distributions (8.2.2022 – 9.2.2022) for the TROPOS reference MPSS 1 (black) and the Lund Dual-MPSS system (red). The shaded area represents the 10% error margin of the TROPOS Reference MPSS No.1. Flow corrections, multiple charge correction and diffusion loss corrections are included. Same as figure 06 with linear y-axis.



**Figure 09:** Correlation of the integrated particle number concentration  $N_{10-800}$  of the MPSS of the Lund Dual-MPSS and the number concentration measured by the total CPC 3750 SN 3750200901.



**Figure 10:** Correlation of the integrated particle number concentration  $N_{10-800}$  of the Reference MPSS 1 and the number concentration measured by the total CPC 3750 SN 3750200901.